

CERTIFICATE OF ANALYSIS

PRODUCT TM1958
M MOD GLUTAMATE D/S 10ML (TRAY 120)

LOT NUMBER 1284775

EXPIRY DATE 2014.12.23

RELEASE DATE 2013.12.25

Physical Testing **Results** **Specification***Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour)	Pass	Purple and durham present
Clarity	Pass	Translucent
pH	6.6	6.6 - 6.8

Microbiological Testing **Result** **Specification** **Method**

Test E.coli MVQC 0014	Pass	Turbid, gas production +ve, acid +ve	QC27
Test E.aerogenes MVQC 0013	Pass	Turbid, gas production +ve, acid +ve	QC27
Test P. aeruginosa MVQC 0023	Pass	Turbid, gas production -ve, acid -ve	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Released by: Rina Pattullo



.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY:
Thermo Fisher Scientific, Microbiology Division
20 Dalglish Street, Thebarton, South Australia 5031
Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054
Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au