



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PO5310A
BRILLIANCE (TM) MRSA 2 AGAR

LOT NUMBER 1285537

EXPIRY DATE 2013.02.07

TYPICAL FORMULATION

	grams per litre
Peptone mix	18.00000
Carbohydrate	4.00000
Salt	5.00000
Kaolin	8.00000
Chromogenic Mix	2.20000
Antibiotic cocktail	20.00000 ML
Agar	13.00000

General Characteristics	Results	Specification
Colour	Conforms	Light ivory
Appearance	Conforms	Opaque
pH	7.4	7.1 - 7.5
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC@33591	66	69	1-2 mm blue colonies
Staphylococcus aureus ATCC@29213	5.6E+04	0	Complete inhibition (<= 10 cfu)
Pseudomonas aeruginosa ATCC@27853	3.2E+04	0	Complete inhibition (<= 10 cfu)
Proteus mirabilis ATCC@29906	9.0E+04	0	Complete inhibition (<= 10 cfu)
Bacillus licheniformis ATCC@14580	<10000	Conforms	Growth of small, rose colonies

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature