



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT **BO0194R**
 T.B.X. AGAR
 10 X 200ML IN 250ML SIROP

LOT NUMBER 1294249

EXPIRY DATE 2015.03.20

Delivery/Customer information

Date Printed

2014.06.05

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw 2-3
pH (25°C)	7.3	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Escherichia coli ATCC®25922	18	24	Blue/green cols
Klebsiella pneumoniae ATCC®29665	36	31	Straw cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

Ian Snowball
Product Performance Manager, G&M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth. PH1 3FX