



Certificate of Analysis

Novex[®] 6% TBE Gel

Product No. EC6265BOX
Lot No. 13020763
Date of Manufacture 07FEB13
Expiration Date 30MAY13

Testing Conditions

Gels were run using Novex[®] TBE Running Buffer and DNA markers. The gels were electrophoresed at 100V until the dye front reached the bottom of the gel.

Migration

2652 bp fragment migrated to 0.150 ± 0.050 Rf, 500 bp fragment to 0.430 ± 0.050 Rf, and 75 bp fragment to 0.865 ± 0.050 Rf.

Straightness

Across the gel, the migration of a given DNA fragment did not vary more than 3% of the length of the gel.

Curvature

In each of the two outermost lanes, the migration of a given DNA fragment did not vary more than 2% of the length of the gel.

Appearance

Gels run were free of swirls, bubbles, and debris. Bands were sharp and flat. There was no DNA degradation.

Overall Result

PASS

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Rick Fletcher
Director, Quality
Issued on 07 Feb 2013