



Certificate of Analysis

Novex[®] 18% Tris–Glycine Gel

Product No. EC6505BOX

Lot No. 13110141

Date of Manufacture 01Nov2013

Expiration Date 13Dec2013

Testing Conditions

Gels were run using Novex[®] Tris–Glycine SDS Running Buffer and unstained protein markers. The gels were electrophoresed at 225V until the dye front reached the bottom of the gel.

Migration

Reduced Myosin migrated to 0.160 ± 0.050 Rf, reduced Trypsin Inhibitor to 0.535 ± 0.050 Rf, and reduced Aprotinin to 0.765 ± 0.050 Rf.

Straightness

Across the gel, the migration of a given protein did not vary more than 3% of the length of the gel.

Curvature

In each of the two outermost lanes, the migration of a given protein did not vary more than 2% of the length of the gel.

Appearance

Gels run were free of swirls, bubbles, and debris. Bands were sharp and flat.

Overall Result

PASS

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Rick Fletcher
Director, Quality
Issued on 02 Nov 2013