

CERTIFICATE OF ANALYSIS

PRODUCT TM1220T5
MALONATE BROTH (5ML)

LOT NUMBER 1331071


EXPIRY DATE 02.08.2013

Physical Characteristics	Results	Specification
Appearance	Clear, green coloured solution	Clear, green coloured solution
pH	6.90	6.50 - 6.90

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification	Quantitative Tests
O2, 35°C, 18-48hr			
<i>Enterobacter aerogenes</i> ATCC®13048	Positive, media blue		
<i>Salmonella typhimurium</i> ATCC®14028	Negative, media green		

Released By



Gan May Ping

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com