



# CERTIFICATE OF ANALYSIS

Page 1 of 1

**PRODUCT** PO5305A  
BRILLIANCE (TM) CAMPYCOUNT AGAR

**LOT NUMBER** 1343665

**EXPIRY DATE** 2013.07.05

## TYPICAL FORMULATION

grams per litre

	grams per litre
Salt Mixture	18.90000
Amino acid mix	3.70000
Vancomycin	0.01000
Anisomycin	0.00980
Cefoperazone	0.02600
Agar	12.00000

General Characteristics	Results	Specification
Colour	Conforms	Signal white
Appearance	Conforms	Transparent
pH	7.5	7.4 - 7.8
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Campylobact. jejuni ssp.jejuni ATCC@33291	86	47	0,5 -2mm maroon colonies
Escherichia coli ATCC@25922	1.8E+04	0	Complete inhibition (<= 10 cfu)
Candida albicans ATCC@10231	5.1E+04	0	Complete inhibition (<= 10 cfu)
Campylobacter coli ATCC@43478	<10000	Conforms	1 - 2mm maroon colonies
Staphylococcus aureus ATCC@25923	>100,000	Conforms	No growth

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature