

VP-SFM
Virus Production
Serum Free Medium 1X
+ 10 ng/mL EGF
- L-Glutamine

Catalog Number: 11681
Lot Number: 1378256

Storage Temperature: 2 to 8C
Storage Instructions: Store in the dark
Expiration Date: 2014-04

For Research or Further Manufacturing Use. CAUTION: Not intended for human or animal diagnostic or therapeutic uses.

TEST	SPECIFICATION	RESULT	UNITS
¹ Endotoxin Testing <i>Tested by manual gel clot method.</i>	>=0 to <=3	0.120	EU/mL
² Osmolality	>=290 to <=350	321	mOsm/kg
³ Performance Assay	>=80 to <=200	139	%
⁴ pH	>=7.0 to <=7.6	7.4	
⁵ Sterility Testing	Negative	Negative	

Maria J. Micale

Quality Systems Department

Date: 15-May-2013

REFERENCES:

- 1 Current United States Pharmacopeia, <85> Bacterial Endotoxins Test.
- 2 Life Technologies Specifications.
- 3 Life Technologies Specifications. Cell line used: African Green Monkey Kidney Cell (VERO), ATCC No. CCL-81. Two 25cm² flasks per control/test media are inoculated with 3 x 10⁵ cells/flask and incubated for four days. The cultures are then harvested, counted, and two new flasks containing fresh growth medium are reinoculated with 3 x 10⁵ cells/flask. This procedure is repeated for a total of three passages. Test lots must be greater than or equal to 80% of control to be acceptable.
- 4 Life Technologies Specifications.
- 5 Current Edition of USP, Life Technologies Modified.

