

# CERTIFICATE OF ANALYSIS

**PRODUCT** PP1297P90  
SENSITEST AGAR + HORSE BLOOD

**LOT NUMBER** 1396517

**EXPIRY DATE** 30.10.2013

Physical Characteristics	Results	Specification
Appearance	Red, opaque gel	Red, opaque gel
pH	7.60	7.20 - 7.60

**Sterility:** Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
<b>O2, 35°C, 24hr</b>		
<i>Staphylococcus aureus</i> ATCC®25923	1mm white colonies	
<i>Escherichia coli</i> ATCC®25922	1-2mm cream colonies	
<i>Pseudomonas aeruginosa</i> ATCC®27853	1-2mm cream colonies	
<i>Enterococcus faecalis</i> ATCC®29212	0.5-1.0mm colourless colonies	

**Released By**



**Nurliyana binti Nazri**

**MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)**

**(Formerly known as Biomedica Laboratories Sdn Bhd)**

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: [oxid.myqcbml@thermofisher.com](mailto:oxid.myqcbml@thermofisher.com)