

Certificate of Analysis

GeneAmp® 10X PCR Buffer I, 2 x 75 mL

Part No.: **4379876**Lot No.: **1401042**

Expiration Date: 30 Apr 2015

COMPONENT	TEST	RESULT
10X PCR BUFFER I Each Bulk Lot of 10X PCR Buffer I is subjected to following tests:	At least 1.5 X 10 ⁵ fold amplification of Lambda DNA after 25 cycles Method: Amplification of Lambda DNA	Pass
	One PCR Product band at 500 base pairs is visualized on a gel Method: Amplification of Lambda DNA	Pass
	Magnesium Concentration: Specification: 14 – 16 mM	Pass

For Research Use Only. Not for human or animal therapeutic or diagnostic use.

ISO 9001:2008 ISO 13485:2003 REGISTERED Austin, TX

Quality Assurance Manager

To obtain a Certificate of Analysis on-line go to www.lifetechnologies.com or email us at cofarequest@lifetech.com