



Certificate of Analysis

Platinum® Quantitative PCR SuperMix–UDG

Product No. 11730–017

Lot No. 1402306

Date of Manufacture 31Jul2013

Product Qualification

Specification: This product is tested functionally in qPCR using genomic DNA. Kinetic analysis must demonstrate a linear dose response with decreasing target concentration and detection from 10 pg human genomic DNA. Components are also tested for the absence of DNase, RNase, and contaminating exonuclease activities.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

A handwritten signature in black ink, appearing to read "Rick Fletcher".

Rick Fletcher
Director, Quality
Issued on 10 Jul 2013