

Certificate of Analysis



GlobalFiler® Express PCR Amplification Kit 1

Product No. **A24602**
Lot No. **1405001**
Date of Manufacture **2014-05-14**
Expiration Date **2015-11-07**

KIT CONFIGURATION

PART NO.	COMPONENT DESCRIPTION	LOT NO.	EXPIRATION DATE
100023653	GlobalFiler® Express PCR Reagents (Box)	1404001	2015-11-07
100023643	GlobalFiler® Express Master Mix (Btl)	1404001	N/A
100023669	GlobalFiler® Express Primer Set (Btl)	1404001	N/A
100023652	Master Mix additive (Tube)	1404001	N/A
100023663	GlobalFiler® Express Allelic Ladder (Box)	1404001	2016-05-06
100023663	GlobalFiler® Express Allelic Ladder (Tube)	1404001	N/A

TEST	SPECIFICATION	RESULT
Functional QC		
QC Method: Samples are amplified and results are compared to a previously approved lot. All analyses are performed using standard operating conditions according to the recommendations in the GlobalFiler® Express Kit User's Guide.		
Human DNA Contamination	Negative Control amplification produces zero (0) human DNA peak.	Pass
Genotype	Four nanograms of Control DNA 007 produces the correct genotype for all 24 loci, including 21 autosomal STR markers, DYS391, Y-indel and the Amelogenin gender marker.	Pass

For Forensic or Paternity Use Only.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Warrington, UK

Quality Assurance
Issued 15May2014

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
www.appliedbiosystems.com
For inquiries, contact us at cofarequests@lifetech.com.

Doc: 100024187 Rev A