

CERTIFICATE OF ANALYSIS

PRODUCT PP1353P90
VANCOMYCIN RESISTANT ENTEROCOCCUS AGAR

LOT NUMBER 1415313

EXPIRY DATE 13.12.2013

Physical Characteristics	Results	Specification
Appearance	Straw green coloured gel	Straw green coloured gel
pH	6.98	6.80 - 7.20

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
O2, 35°C, 24-48hr		
<i>Enterococcus faecalis</i> ATCC®51299	1.0mm colourless cols w black background	
<i>Enterococcus faecalis</i> ATCC®29212	Complete inhibition	
<i>Escherichia coli</i> ATCC®25922	Partial inhibition	

Released By



Nurliyana binti Nazri