

CERTIFICATE OF ANALYSIS

PRODUCT PP1150P90
DNASE AGAR

LOT NUMBER 1417961

EXPIRY DATE 19.12.2013

Physical Characteristics	Results	Specification
Appearance	Light to medium amber coloured agar	Light to medium amber coloured agar
pH	7.17	7.10 - 7.50

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification	Quantitative Tests
---------------------------------	---------	---------------	--------------------

O2, 35°C, 24-48hr

Staphylococcus aureus ATCC®25923

Dnase (+), clear zone around colonies

Staph. epidermidis ATCC®12228

Dnase (-), no zone around colonies

Klebsiella pneumoniae ATCC®13883

Dnase (-), no zone around colonies

Released By



Nurliyana binti Nazri

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com