



Certificate of Analysis

Purelink™ HiPure Precipitators 10 Reactions

Product No. K210021
Lot No. 1441328
Date of Manufacture 30Sep2013

Filter Performance

Average Flow Rate : 145mL/min@30 psi
Bubble Point : >= 15psi
Maximum Operating Pressure : 60psi

DNA Recovery

200µg Plasmid DNA was processed over the filter device as described in the product manual, and a minimum recovery of 70% Plasmid DNA obtained.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

A handwritten signature in black ink that reads "Rick Fletcher".

Rick Fletcher
Director, Quality
Issued on 06 Sep 2013