

CERTIFICATE OF ANALYSIS

PRODUCT PP1405P90
CHROMAGAR CANDIDA AGAR (22ML)

LOT NUMBER 1448218

EXPIRY DATE 07.02.2014

Physical Characteristics	Results	Specification
Appearance	Straw, translucent gel	Straw, translucent gel
pH	6.04	5.90 - 6.30

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification	Quantitative Tests
O2, 35°C, 48hr			
<i>Candida albicans</i> ATCC®10231	1-2mm green colonies		
<i>Staphylococcus aureus</i> ATCC®25923	Complete inhibition		
<i>Escherichia coli</i> ATCC®25922	Complete inhibition		

Released By



Gan May Ping