

Certificate of Analysis

Novex® 10% Tris-Glycine Midi Gel

Product No. WT0102BX10

Lot No. 15021181

Date of Manufacture 11Feb2015 Expiration Date 03Jun2015

Testing Conditions

Gels were run using Novex® Tris-Glycine SDS Running Buffer and unstained protein markers. The gels were electrophoresed at 225V until the dye front reached the bottom of the gel.

Migration

Reduced Myosin migrated to 0.235 ± 0.050 Rf, reduced BSA to 0.500 ± 0.050 Rf, and reduced Carbonic Anhydrase to 0.915 ± 0.050 Rf.

Straightness

Across the gel, the migration of a given protein did not vary more than 5% of the length of the gel.

Curvature

In each of the two outermost lanes, the migration of a given protein did not vary more than 2% of the length of the gel.

Appearance

Gels run were free of swirls, bubbles, and debris. Bands were sharp and flat.

Overall Result

PASS

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies 5791 Van Allen Way Carlsbad, CA, USA 92008 www.lifetechnologies.com

For inquiries, contact us at $\underline{cofarequests@lifetech.com}$

Cord R. Bleebly

Carol Bleakley Quality Leader Issued on 11 Feb 2015