



Certificate of Analysis

Hu JAK 2[pYpY1007/1008]

Product No. KHO5621
Lot No. 1531315A
Expiration Date 30Jun2015

<u>Part Number</u>	<u>Lot Number</u>	<u>Description</u>	<u>Part Number</u>	<u>Lot Number</u>	<u>Description</u>
SD237	A10924	Standard	PP237	1450825	Ab Coated Plates
DS36A	874163A	Standard Diluent	WB01	As Supplied	Wash Buffer
UD237	1531317	Biotin Conjugate	SB01	72852680F	Chromogen
SH237	1531316	Anti-Rabbit IgG HRP	SS03	As Supplied	Stop Solution
DV10I	1312227A	HRP Diluent			

Parameter	Standard Conc. pg/mL	Avg. O.D.
Standard 1	100.0	2.637
Standard 2	50.0	1.655
Standard 3	25.0	0.917
Standard 4	12.5	0.508
Standard 5	6.3	0.286
Standard 6	3.1	0.173
Standard 7	1.6	0.120
Standard 8	0.0	0.067

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Michael Grossman
Sr. Manager, GOS – Quality
Issued on 20 Feb 2014