



## Certificate of Quality

**Product Name:** NLA - MRS Agar 100g each  
**Lot Number:** 155837

**Product Number:** R454051  
**Expiration Date:** 2016-02-29  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Lactobacillus casei	9595	Good Growth
Lactobacillus acidophilus	33200	Good Growth
Lactobacillus acidophilus	4356	Good Growth
Lactobacillus fermentum	9338	Good Growth

Visual Driform: Light beige, moist powder

pH: Upon rehydration per manufacturer specifications, media has a pH of 6.2 +/- 0.2

This lot meets Remel's criteria for approval as defined for this product. It should be stable to the expiration date on the label when stored according to package specifications.

Signed

A handwritten signature in black ink that reads "Susie Waddell". The signature is written in a cursive, flowing style.

Quality Assurance Supervisor