



# Certificate of Analysis

AccuPrime™ Taq DNA Polymerase

Product No. 12339-016

Lot No. 1596013

Date of Manufacture 01May2014

Expiration Date 31May2016

---

## PCR Funtional Assay

Specification: Negative control (no template) reactions must be free of visible product.

Result: Meets specification.

Specification: Reactions must show the amplification of a 626 bp band from 100 ng of human genomic DNA and a 1989 bp band from 100 pg of pBR322 DNA. An increased specificity and yield of PCR product from human genomic DNA must be observed in reactions containing AccuPrime™ protein versus reactions without.

Result: Meets specifications.

---

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.lifetechnologies.com](http://www.lifetechnologies.com)  
For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com)

---

Carol Bleakley  
Quality Leader  
Issued on 09 May 2014