

Certificate of Analysis

Novex® 12% Tris–Glycine Gel

Product No. EC600

Lot No. 16081243

Date of Manufacture 12Aug2016

Expiration Date 04Nov2016

This certificate of analysis applies to the following catalog numbers:

<u>Catalog No.</u>	<u>Description</u>	<u>Pack Size</u>	<u>Catalog No.</u>	<u>Description</u>	<u>Pack Size</u>
EC6001BOX	1.0mm 1–well	10 gels/box	EC6006BOX	1.0mm 2D–well	10 gels/box
EC6005BOX	1.0mm 10–well	10 gels/box	EC6008BOX	1.5mm 10–well	10 gels/box
EC60052BOX	1.0mm 12–well	10 gels/box	EC60085BOX	1.5mm 15–well	10 gels/box
EC60055BOX	1.0mm 15–well	10 gels/box			

Testing Conditions

Gels were run using Novex® Tris–Glycine SDS Running Buffer and unstained protein markers. The gels were electrophoresed at 225V until the dye front reached the bottom of the gel.

Migration

Reduced Myosin migrated to 0.200 ± 0.050 Rf, reduced LDH to 0.630 ± 0.050 Rf, and reduced Trypsin Inhibitor to 0.890 ± 0.050 Rf.

Straightness

Across the gel, the migration of a given protein did not vary more than 3% of the length of the gel.

Curvature

In each of the two outermost lanes, the migration of a given protein did not vary more than 2% of the length of the gel.

Appearance


Gels run were free of swirls, bubbles, and debris. Bands were sharp and flat.

Overall Result

Meets Specification

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Anita Targosz
QA Associate Director
Issued on 12 Aug 2016