



Certificate of Quality

Product Name: Anaer Reducible LKV Blood Agar 15/PK
Lot Number: 166278

Product Number: R01069
Expiration Date: 2012-05-15
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides fragilis	25285	Good Growth
Prevotella melaninogenica	25845	Good Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth

Physical Characteristics:

Appearance: Brown to Brown-Red, Clear Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor