



## Certificate of Quality

**Product Name:** XLT4/Bril Green w/Novo Biplate 100/PK  
**Lot Number:** 169482

**Product Number:** R110506  
**Expiration Date:** 2012-05-22  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
<b>BG with Novo:</b>		
Salmonella typhimurium	14028	Pink Colonies w/ Pink Zone
Proteus mirabilis	12453	No Growth
<b>XLT4:</b>		
Salmonella typhimurium	14028	Red Colonies w/ Black Centers
Proteus mirabilis	12453	No Growth
Citrobacter freundii	33128	Yellow Colonies

### **Physical Characteristics:**

Appearance: Brilliant Green Side: Green, Clear Agar  
XLT4 Side: Red-Orange, Clear Agar  
pH @ 20-25°C: N/A

### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor