

# Certificate of Analysis

## Human PAI-1 Magnetic Bead Kit

Product No. LHC0221M  
Lot No. 1699551A  
Date of Manufacture 02Sep2015  
Expiration Date 30Sep2017

---

Part Number	Lot Number	Description
SM320	1394809	Standard
LM143M	1699552	Antibody Bead Concentrate
BN143	1699553	Detection Antibody Concentrate
LHB0004M	1699904	Buffer kit use in Final Release Test


Parameter	Standard Conc. pg/mL	Avg. MFI
Standard 1	140000	17439
Standard 2	46667	13991
Standard 3	15556	7329
Standard 4	5185	2626
Standard 5	1728	688
Standard 6	576	160
Standard 7	192	52
Background	0	17

---

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)

For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com)



---

Stephen McKinley  
Senior QAQC Manager  
Issued on 19 Oct 2015