

Certificate of Analysis

ZOOM[®] Strips, pH 3–10NL

Product No. ZM0011
Lot No. 170119
Date of Manufacture 19Jan2017

Visual Inspection

All Strips sampled are free of visual flaws.
All information (lot #,s trip range, serialization) printed on strips is correct

Migration

Five bands on the strip are measured after first dimension run is complete.


Band A – 0.17 ± 0.06 Rf
Band B – 0.32 ± 0.06 Rf
Band C – 0.46 ± 0.06 Rf
Band D – 0.63 ± 0.06 Rf
Band E – 0.89 ± 0.06 Rf

Overall Result

PASS

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Anita Targosz
QA Associate Director
Issued on 20 Jan 2017