



## Certificate of Quality

**Product Name:** Columbia Agar MLT 10/PK  
**Lot Number:** 231311

**Product Number:** R110165  
**Expiration Date:** 2012-10-15  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Clostridium sporogenes	11437	Pass

A pass result indicates organism produced the expected colony size, number, and morphology, indicative of the challenge organism and level for the medium formulation, within the required time period, e.g., two days for bacteria at 30-35C. All inoculations are performed with <100 CFU per challenge. The results meet or exceed USP Growth Promotion requirements.

### Physical Characteristics

Appearance: Light to Medium Amber, Clear Agar w/ Slight Haze  
pH @ 20-25C: 7.3 +/- .2

### Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum 72 days at 30-35°C.

Certificate results were obtained at time of release. End users are encouraged to assure product meets internal standards at time of use.

Signed

A handwritten signature in black ink that reads "Susie Weddle".

Quality Assurance Supervisor