



Certificate of Quality

Product Name: Baird Parker Agar 10/PK
Lot Number: 239254

Product Number: R01108
Expiration Date: 2012-10-25
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|----------------------------|-------------|---------------------------------|
| Staphylococcus aureus | 25923 | Black Colonies, Opaque Zone |
| Escherichia coli | 25922 | No Growth |
| Staphylococcus epidermidis | 12228 | Inhibition, Partial to Complete |

Physical Characteristics:

Appearance: Yellow, Semitransparent Agar
pH @ 20-25°C: 7.0 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor