

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

## Certificate of Analysis

Product Name: Coag Control Level 2 (Abnormal)-Orig:USA Product Number: 250596 Lot Number: 239407 Expiration Date: 2015-05-31 Storage: 2 - 8°C

Specification	Specification Limit	Result
VAN Number	Report	187001
Date of Manufacture	Report	23-MAY-12

For in-vitro diagnostic use.

\* Calculated from recovered values. Actual values recovered in the user laboratory will depend on reagent lot number, instrument and technique in use. Each laboratory should establish ranges using instrument, reagent and technique in place.

- All individual donor units comprising this lot of Coagulation Control Plasma have been tested and found nonreactive for HBsAg and negative for antibodies to HIV and HCV. As with all blood products, this product should be handled as potentially infectious biological material.

- Handling: Reconstitute with 1.0 mL high purity water. Swirl gently and let stand undisturbed for 15 minutes at room temperature.

- Stability: Unopened vials are stable until the expiration date indicated above when stored at 2-8 °C. Reconstituted control is stable for 8 hours when stored in a capped vial at 2-8 °C.

 $\mathcal{A}$ 

Quality Representative

Date: 2012-10-11