



Certificate of Quality

Product Name: Contact I Mycological w/Lec,PS80 20/PK
Lot Number: 290816

Product Number: R111030
Expiration Date: 2013-02-21
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Candida albicans	10231	Good Growth
Cryptococcus neoformans	34877	Good Growth
Trichophyton mentagrophytes	9533	Good Growth

Physical Characteristics:

Appearance: Medium Amber, Clear Agar
pH: 4.7 +/- 0.2

Certificate results were obtained at time of release. End users are encouraged to assure product meets internal standards at time of use.

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor