



Certificate of Quality

Product Name: TSA 10kg each
Lot Number: 331288

Product Number: R455006
Expiration Date: 2017-05-31
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria	
		Without Blood	With Blood
Bacillus subtilis	6633	Good Growth	
Escherichia coli	8739	Good Growth	
Pseudomonas aeruginosa	9027	Good Growth	
Staphylococcus aureus	6538	Good Growth	
Streptococcus pneumoniae	6305	Growth	Good Growth, Alpha Hemolysis
Streptococcus agalactiae	12386	Good Growth	Good Growth, Beta Hemolysis, CAMP Test (+)
Streptococcus pyogenes	19615	Good Growth	Good Growth, Beta Hemolysis, CAMP Test (-)
Streptococcus pyogenes	12377		Good Growth, Beta Hemolysis
Streptococcus pyogenes	12384	Good Growth	Good Growth, Beta Hemolysis

Visual Driform: Light beige, homogeneous, free-flowing powder

pH: Upon rehydration per manufacturer specifications, media has a pH of 7.3 +/- 0.2

*Prior to dispensing, 5% defibrinated sheep blood is added to the prepared, cooled basal medium.

This lot meets Remel's criteria for approval as defined for this product. It should be stable to the expiration date on the label when stored according to package specifications.

Signed

Quality Assurance Supervisor