



## Certificate of Quality

**Product Name:** RapID ANA II System 20/KT  
**Lot Number:** 392879

**Product Number:** R8311002  
**Expiration Date:** 2014-09-17  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Clostridium sordellii	9714	Acceptable
Bacteroides uniformis	8492	Acceptable
Bacteroides distasonis	8503	Acceptable
Bacteroides fragilis	25285	Acceptable
Bacteroides thetaiotaomicron	29741	Acceptable
Clostridium perfringens	13124	Acceptable
Propionibacterium acnes	6919	Acceptable
Peptostreptococcus anaerobius	27337	Acceptable
Bacteroides vulgatus	8482	Acceptable

Signed

A handwritten signature in black ink, appearing to read "S. Kendall", written in a cursive style.

Quality Control Manager