



Certificate of Quality

Product Name: Inhibitory Mold Agar w/Gent (26ml) 10/PK
Lot Number: 402165

Product Number: R01506
Expiration Date: 2013-12-12
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Candida albicans	10231	Good Growth
Cryptococcus neoformans	34877	Good Growth
Trichophyton mentagrophytes	9533	Good Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth
Pseudomonas aeruginosa	27853	No Growth

Physical Characteristics:

Appearance: Light to Medium Amber, Clear Agar*
pH @ 20-25°C: 6.7 +/- 0.2

*Note: Medium may contain a small quantity of dark particles, which do not impair performance.

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor