



Certificate of Quality

Product Name: MRS Agar 10/PK
Lot Number: 408626

Product Number: R01585
Expiration Date: 2013-12-09
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Lactobacillus johnsonii	33200	Good Growth

Physical Characteristics:	
Appearance:	Medium Amber, Clear Agar
pH @ 20-25°C:	6.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor