



Certificate of Quality

Product Name: ^MN^ 10/CS TSB CAP IV 500ML
Lot Number: 414998

Product Number: R112784
Expiration Date: 2014-04-22
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Inoculum	Performance Criteria
Staphylococcus aureus	6538	Quanti Cult	Growth
Pseudomonas aeruginosa	9027	Quanti Cult	Growth
Escherichia coli	8739	Quanti Cult	Growth
Salmonella species	14028	Quanti Cult	Growth

Media Physical Characteristics:

Appearance: Light to Medium Amber, Clear to Slightly Hazy Broth
pH @ 20-25°C: N/A

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor