



## Certificate of Quality

**Product Name:** M6 (1.5ml) PK/12  
**Lot Number:** 457082

**Product Number:** R12535  
**Expiration Date:** 2015-07-30  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
<b>Maintenance:</b>		
Chlamydia trachomatis	VR-880	Pass
Mycoplasma hominis	23114	Pass
Mycoplasma pneumoniae	15531	Pass
Ureaplasma urealyticum	27618	Pass
Herpes simplex type 2	VR-734	Pass
Cytomegalovirus	VR-977	Pass
Respiratory Syncytial virus	VR-1302	Pass

<b>Toxicity:</b>		
MRC5 Cells	CCL-171	Pass
VERO Cells	CCL-81	Pass
McCoy Cells	RLL-169	Pass

<b>Physical Characteristics:</b>	
Appearance:	Orange-Pink Clear, Broth w/ Clear Beads.
pH @ 20-25°C:	7.1 - 7.5

### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor