



Certificate of Quality

Product Name: U9B Complete (2ml) 20/PK
Lot Number: 468168

Product Number: R20385
Expiration Date: 2014-08-25
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge

Ureaplasma urealyticum
Mycoplasma hominis
Staphylococcus aureus

ATCC

27618
23114
25923

Performance Criteria

Red Color Change, Good Recovery on A-7 Agar
No Color Change, Good Recovery on A-7 Agar
Marked to Complete Inhibition

Physical Characteristics:

Appearance:
pH @ 20-25°C:

Clear, Yellow Broth
N/A

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor