



Certificate of Quality

Product Name: Todd Hewitt Broth w/CNA (LIM)(3ml)100/PK
Lot Number: 483132

Product Number: R064810
Expiration Date: 2015-03-31
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Streptococcus agalactiae	12386	Growth, (+) For Latex Agglutination With Group B Reagent
Escherichia coli	25922	No Growth, (-) For Latex Agglutination With Group B Reagent

Physical Characteristics:
Appearance: Yellow, Clear Broth
pH @ 20-25°C: 7.8 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell". The signature is written in a cursive, flowing style.

Quality Assurance Supervisor