



Certificate of Quality

Product Name: MH Agar (150mm) 40/PK
Lot Number: 491292

Product Number: R04052
Expiration Date: 2014-07-14
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Staphylococcus aureus	25923	All Zones within NCCLS Guidelines
Escherichia coli	25922	All Zones within NCCLS Guidelines
Pseudomonas aeruginosa	27853	All Zones within NCCLS Guidelines
Escherichia coli	35218	All Zones within NCCLS Guidelines

This product has been tested and is compliant with the performance criteria in the NCCLS document, "Performance Standards For Antimicrobial Disk Susceptibility Tests."

Physical Characteristics:

Appearance: Light Amber, Clear Agar
pH @ 20-25°C: 7.3 +/- 0.1

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor