



Certificate of Quality

Product Name: Blood/SXT Biplate 10/PK
Lot Number: 503221

Product Number: R02053
Expiration Date: 2014-07-28
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Blood Agar:		
Streptococcus pyogenes	19615	Good Growth, Beta Hemolysis
Streptococcus agalactiae	12386	Good Growth, Beta Hemolysis
Enterococcus faecalis	29212	Good Growth, Gamma Hemolysis
Streptococcus pneumoniae	6305	Good Growth, Alpha Hemolysis
Listeria monocytogenes	7646	Good Growth
Escherichia coli	25922	Good Growth
Staphylococcus aureus	25923	Good Growth
Moraxella catarrhalis	25238	Good Growth

SXT Blood Agar:		
Streptococcus pyogenes	19615	Growth, Beta Hemolysis
Staphylococcus epidermidis	12228	Inhibition, Partial to Complete
Streptococcus sanguinis	10556	Marked to Complete, Inhibition

Physical Characteristics:	
Appearance:	Blood Agar: Red, Opaque Agar SXT Blood Agar: Red, Opaque Agar
pH @ 20-25°C:	Blood Agar: 7.3 +/- 0.2 SXT Blood Agar: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor