

Fisher Diagnostics a division of Fisher Scientific Company, LLC a part of Thermo Fisher Scientific Inc. 8365 Valley Pike

Middletown, VA (USA) 22645-1905 Tel: 540 869 3200 or 800-528-0494

Certificate of Analysis

Product Name: Clinical Chemistry Acid Wash
OEM Part Number: 06K010020
OEM Lot Number: 56656UN17
Storage: 15 - 30°C
Abbott Commodity Number: 39-6412
Expiration Date: 2018-04-04
Date of Manufacture: 04-JAN-17

Qty Manufactured: 4,140 Kits

Specification	Specification Limit	Result
Appearance	Clear, colorless liquid free of	Complies Pass
	undissolved materials.	
pH (pH units)	0.5 - 1.5	1.1 Pass
Refractive Index (Brix % Scale)	8.0 - 9.0 %	8.4 Pass
Specific Gravity	1.01 - 1.03 g/mL	1.03 Pass

For in-vitro diagnostic use.

This lot meets all current release specifications per quality document MS.MASTER342.

The above product has been manufactured in compliance with Quality Systems Regulations, passed all Thermo Fisher final release specifications, and has met all Abbott acceptance specifications.

This product contains a single component. The component name, lot number, and expiration date matches the kit name, lot number, and expiration date listed on this COA.

Minimum Fill Volume: >= 500 mL Complies Pass

Quality Representative Date: 2017-01-16

PRINTED ON: 2017-01-16