



Quality Assurance Certificate

Code: BC0100M
Product: SIGNAL BLOOD CULTURE SYSTEM 1 X 20
Lot: 623038
Expiry: 31 January 2010
Components:

PHYSICAL CHARACTERISTICS: As appropriate tests on moisture, appearance, pH, clarity and gel strength were satisfactory.

BIOLOGICAL PERFORMANCE: Satisfactory performance of:

Clostridium perfringens	ATCC 13124
Bacillus cereus	ATCC 10876
Bacteroides fragilis	ATCC 25285
Clostridium novyi	ATCC 27606
Escherichia coli	ATCC 25922
Fusobacterium nucleatum	ATCC 10953
Haemophilus influenzae	ATCC 19418
Klebsiella pneumoniae	ATCC 29665
Neisseria meningitidis	ATCC 13077
Pseudomonas aeruginosa	ATCC 27853
Staphylococcus aureus	ATCC 25923
Staphylococcus epidermidis	ATCC 14990
Streptococcus pneumoniae	ATCC 6303
Streptococcus mutans	ATCC 25175
Candida albicans	ATCC 10231
Prevotella bivia	ATCC 29303

Date: 11 November 2013

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.

C. MACBRIDE
Product Performance Manager

C. BOOTH
V.P. Science and Technology



Certificate No. FM09914
BS EN ISO 9001
BS EN ISO 13485

OXOID LIMITED
Wade Road, Basingstoke, Hants RG24 8PW, England
Oxoid@Oxoid.com www.oxoid.com

FDA Reg. No. 8010096