



Certificate of Quality

Product Name: Thio Med w/Dex,Hem,Vit K(10ml) 100/PK
Lot Number: 636141

Product Number: R07180
Expiration Date: 2008-11-19
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	Good Growth
Staphylococcus aureus	25923	Good Growth
Clostridium perfringens	13124	Good Growth
Peptostreptococcus anaerobius	27337	Good Growth
Bacteroides vulgatus	8482	Good Growth

Physical Characteristics:

Appearance: Light to Medium Amber, Clear Broth with Visible Agar
pH @ 20-25°C: 7.0 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor