

Certificate of Quality

Product Name: Haemophilus ID II Triplate 10/PK
Lot Number: 639066

Product Number: R02345
Expiration Date: 2008-04-14
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
QUADRANT I		
Haemophilus influenzae	10211	Growth
Haemophilus influenzae	49247	Growth
Haemophilus parainfluenzae	7901	Growth
Haemophilus parahaemolyticus	10014	Growth, Beta Hemolysis
Haemophilus aphrophilus	19415	Growth

QUADRANT II		
Haemophilus influenzae	10211	No Fluorescence
Haemophilus influenzae	49247	No Fluorescence
Haemophilus parainfluenzae	7901	Orange Fluorescence
Haemophilus parahaemolyticus	10014	Orange Fluorescence
Haemophilus aphrophilus	19415	Orange Fluorescence

QUADRANT III		
Haemophilus influenzae	10211	No Growth
Haemophilus influenzae	49247	No Growth
Haemophilus parainfluenzae	7901	No Growth
Haemophilus parahaemolyticus	10014	No Growth
Haemophilus aphrophilus	19415	Growth

Physical Characteristics:

Appearance:	QUADRANT I: Red, Opaque Agar QUADRANT II: Brown, Opaque Agar QUADRANT III: Light To Medium Amber, Clear Agar
pH @ 20-25°C:	QUADRANT I: 7.4 +/- 0.2 QUADRANT II: 6.9 +/- 0.2 QUADRANT III: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Certificate of Quality

Signed

A handwritten signature in black ink, reading "Susie Waddell". The signature is fluid and cursive, with the first name "Susie" and last name "Waddell" clearly distinguishable.

Quality Assurance Supervisor