

Certificate of Quality

Product Name: Anaer Reducible CNA Bld Agar 15/PK
Lot Number: 642604

Product Number: R01065
Expiration Date: 2008-05-20
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides fragilis	25285	Good Growth
Peptostreptococcus anaerobius	27337	Good Growth
Staphylococcus aureus	25923	Good Growth
Escherichia coli	25922	Inhibited

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor