

Certificate of Quality

Product Name: HE (Hektoen Enteric)/XLD Biplate 10/PK
Lot Number: 653630

Product Number: R02086
Expiration Date: 2008-06-24
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
HE AGAR:		
Salmonella typhimurium	14028	Green Colonies, Black Centers
Shigella flexneri	12022	Green Colonies
Enterococcus faecalis	29212	No Growth
Escherichia coli	25922	Marked to Complete Inhibition

XLD AGAR:		
Salmonella typhimurium	14028	Red Colonies, Black Centers
Shigella flexneri	12022	Red Colonies
Enterococcus faecalis	29212	No Growth
Escherichia coli	25922	Marked to Complete Inhibition

Physical Characteristics:	
Appearance:	HE Agar: Green, Clear Agar XLD Agar: Red-Orange, Clear Agar
pH @ 20-25°C:	HE Agar: 7.6 +/- 0.2 XLD Agar: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor