

Certificate of Quality

Product Name: D/E Neut Broth (5ml) 100/PK
Lot Number: 657932

Product Number: R060770
Expiration Date: 2009-04-29
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	Good Growth, Yellow
Pseudomonas aeruginosa	27853	Good Growth
Staphylococcus aureus	25923	Good Growth, Yellow

Physical Characteristics:

Appearance: Purple-Gray, Semi-Transparent Broth
pH @ 20-25°C: 7.6 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor