



## Certificate of Quality

**Product Name:** Middlebrook 7H11 Agar (slant) 100/PK  
**Lot Number:** 670735

**Product Number:** R08570  
**Expiration Date:** 2009-06-11  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Mycobacterium fortuitum	6841	Good Growth
Mycobacterium intracellulare	13950	Good Growth
Mycobacterium kansasii	12478	Good Growth
Mycobacterium scrofulaceum	19981	Good Growth
Mycobacterium tuberculosis	25177	Good Growth

### **Physical Characteristics:**

Appearance: Faint Green, Clear Agar Slant  
pH @ 20-25°C: 6.6 +/- 0.2

### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor