



Certificate of Quality

Product Name: NLA - EC Medium 100g each
Lot Number: 676103

Product Number: R453301
Expiration Date: 2012-11-27
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	Good Growth, Gas Production
Enterobacter aerogenes	13048	Good Growth
Enterococcus faecalis	29212	Inhibited
Bacillus subtilis	6633	Inhibited

Visual Driform: Light beige, homogeneous, free-flowing powder

pH: Upon rehydration per manufacturer specifications, media has a pH of 6.9 +/- 0.2

This lot meets Remel's criteria for approval as defined for this product. It should be stable to the expiration date on the label when stored according to package specifications.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor