

# Certificate of Quality

**Product Name:** SXT Blood Agar 15/PK  
**Lot Number:** 678746

**Product Number:** R01861  
**Expiration Date:** 2008-08-28  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Streptococcus pyogenes	19615	Growth, Beta Hemolysis
Staphylococcus epidermidis	12228	Inhibition (Partial to Complete)
Streptococcus sanguinis	10556	Inhibition (Partial to Complete)

#### **Physical Characteristics:**

Appearance: Red, Opaque Agar  
pH @ 20-25°C: 7.3 +/- 0.2

#### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor