

Certificate of Quality

Product Name: MacConkey Agar 15/PK
Lot Number: 692022

Product Number: R01551
Expiration Date: 2008-10-27
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Proteus mirabilis	12453	Clear Colonies, No Swarming
Proteus vulgaris	6380	Clear Colonies, No Swarming
Escherichia coli	25922	Pink Colonies
Salmonella typhimurium	14028	Clear Colonies
Staphylococcus aureus	25923	No Growth
Enterococcus faecalis	29212	No Growth

Physical Characteristics:

Appearance: Pink, Clear Agar
pH @ 20-25°C: 7.1 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor